

CALIBRATION THERMISTOR TEST CAPS

- Provides assurance that thermometer readings are accurate
- Supplied with a UKAS Certificate of Calibration

These thermistor test caps are suitable for checking the accuracy of the Therma 20, Therma 22 or any equivalent thermistor thermometer.

Simply plug in the desired test cap and the display on the thermometer should show the same temperature as the certified value.

Each test cap is supplied with a UKAS Certificate of Calibration with a guaranteed uncertainty of $\pm 0.1^\circ\text{C}$.



Order code	Description	
286-001	Thermistor test cap -18°C	4099162
286-002	Thermistor test cap 0°C	4099163
286-003	Thermistor test cap 3°C	4099164
286-004	Thermistor test cap 70°C	4099165
286-005	Thermistor test cap 100°C	4099166

UKAS Certificate of Calibration included

CALIBRATION PT100 TEST CAPS

- Validates the accuracy of PT100 thermometers
- Supplied with a UKAS Certificate of Calibration

These PT100 test caps are suitable for checking the accuracy of the Precision PT100 thermometer or any platinum resistance thermometer fitted with a Binder connector.

Simply plug in the test cap and the display on the thermometer should show the same temperature as the certified value.

Each test cap is supplied with a UKAS Certificate of Calibration with a guaranteed uncertainty of $\pm 0.1^\circ\text{C}$.



Order code	Description	
282-001	PT100 test cap -18°C	4099151
282-002	PT100 test cap 0°C	4099152
282-003	PT100 test cap 3°C	4099153
282-004	PT100 test cap 70°C	4099154
282-005	PT100 test cap 100°C	4099155

UKAS Certificate of Calibration included

CALIBRATION EQUIPMENT